

To the customers and users of heated Dräger breathing circuits

May 2020

Important safety notice!!!

Combinations of breathing circuits with humidifiers

Product affected:

Only heated Dräger breathing circuits are affected

Dear Sir or Madam,

As part of our global product monitoring, Dräger has become aware of a case in which a VentStar Helix heated breathing circuit was used together with a Fisher & Paykel MR810 humidifier. This product combination is not approved. The humidifier used did not detect previously unexplained damage to the heating wire of the VentStar Helix Heated. This led to the ignition of the breathing gas and the affected patient suffered burns.

If Dräger breathing circuits are used with humidifiers other than the Fisher & Paykel MR850 humidifier indicated in our instructions for use, patients and/or users can be injured if an error occurs.

The Fisher & Paykel MR850 humidifier has control functions that would have disconnected the power supply and generated an alarm if the above error had occurred.

Make sure that all heated Dräger breathing circuits are only used together with the Fisher & Paykel MR850 humidifier indicated in the instructions for use

Please inform all relevant users in your hospital.

We regret any inconvenience this information may cause but consider it necessary as a preventive measure to increase patient and user safety.



Page 2 / 2

With many thanks in advance for your support.

Yours faithfully,

Markus Kämer

Senior Product Manager BU Hospital Consumables & Accessories

Drägerwerk AG & Co. KGaA

Oliver Möller

Deputy Safety Officer for Medical Devices Post Market Surveillance Drägerwerk AG & Co. KGaA